

# ◀ HEALTH & SAFETY ALERT ▶



## Fluke Electrical Equipment Recall

This safety alert was produced by our colleagues at Carillion. It relates to technical compliance of Fluke electrical equipment.

Fluke is voluntarily recalling certain 373, 374, 375 and 376 Digital Clamp Meters that were manufactured between 1<sup>st</sup> September 2010 and 31<sup>st</sup> October 2012. The models and serial numbers affected are:

Model Family	Manufacture Dates	Starting Serial No.	Ending Serial No.
Fluke-376	Sep 2010 - Feb 2011	14270001	15909999
Fluke-375	Sep 2010 - Mar 2011	14270001	16079999
Fluke-374	Oct 2010 - Feb 2011	14270001	16379999
Fluke-373	Oct 2010 - Oct 2012	14270001	21950000



Any serial number preceded or followed by the letter "R" is not affected by this recall notice. The image above shows where to find the model and serial number of the effected clamp meters.

The printed circuit assembly may not be properly fastened to the test lead input jack. This may result in inaccurate voltage readings, including a low or no voltage reading on a circuit energized with a hazardous voltage, presenting a shock, electrocution or thermal burn hazard.

If you own one of these clamp meters, please stop using them Immediately even if you have not experienced problems and send it back to Fluke for repair (please only send the clamp meter – any test leads/boxes/manuals that are sent with the unit will NOT be returned.)

Call +44-20 7942 0700 and request an email with a printable, pre-paid and pre-addressed shipping label.

Further Information: <http://www.fluke.com/fluke/usen/support/safety/37X-Recall.htm>